



INDIRA GANDHI UNIVERSITY, MEERPUR, REWARI

(A State University Established under Haryana Act No. 29 of 2013)

No. IGU/Botany/2020/ 215- 230

Dated: 20.10.2020

To

Subject: - Sealed quotation for Laminar flow Due on 06.11.2020.

Dear Sir/Madam,

Please quote for the article mentioned below so as to reach the undersigned by the Department of Botany. The quality, make, packaging and other particulars of each item should be stated.

WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

1. The quotations are to be sent in a sealed envelope marked QUOTATION DUE ON 06-11-2020 and this envelope may be put in another envelope addressed to the undersigned.
2. The rates of S.T/C.S.T./GST, EXCISE, CUSTOM and another taxes chargeable must be specified.
3. The rates of Insurance, if any, should be specified. The firm will be required to submit original Payee Receipt along with the bill.
4. Please state the time period within which the items will be supplied.
5. GUARANTEE PERIOD OF THE INSTRUMENT MUST BE MENTIONED.
6. Discount if any, is to be given with the quotation
7. The rates quoted should be FOR destination Meerpur, Rewari.
8. Please state, if you are the original manufacturer/sole distributor of the items quoted by you.

Sl. No.	Name of item	Quantity	Make	Model Cat. No.	Other Specification
1.	Laminar flow	two	Atlantis		<ol style="list-style-type: none"> 1. Vertical down wards flow of Ultra Clean Air at 90 + 20 fpm. 2. Open front 3. Side Panels of both sides laminated ply 4. Efficiency of Ultra Clean Air down to 0.3 micron (DOP) dust particles is 99.97%. 5. Conform to Air Cleanliness tests as per Article 5.1 of U.S. Federal Standard 209 (Class 100) latest version 6. Total Height of the Vertical Flow Bench - 7 ft. (Approx.) 7. Continuous use Blower Assembly with 1/4 H.P. Motor(s), 50 Hz 1 ph to deliver air at 550 8. CFM (Approx.) 50 mm.WG Pressure 9. Cabinet: All ply wood inside painted with synthetic Enamel Paint and outside with white Laminate to match the elegance and decor of your research lab. 10. Work Table: Commercial ply board with white surface affixed on it, easily cleanable, alternatively SS 11. Illumination 230 VAC 50 Hz Fluorescent tube lights 12. Controls: Independent 16 A switches

Thanking you,

Yours sincerely,


 Chairperson

CC to:-

A copy of the above is forwarded to incharge University website for uploading the same on University website.