



INDIRA GANDHI UNIVERSITY, MEERPUR, REWARI

(A State University Established under Haryana Act No. 29 of 2013)

No. IGU/BOTANY/2021/ 453-467

Dated: 18.01.2021

To

Subject: - Sealed quotation for Orbital Shaker incubator (LE-TT-A) due on 03.02.2021.

Dear Sir/Madam,

Please quote for the article mentioned below so as to reach the undersigned by 03.02.2021 up to 05:00 pm in the office of Chairperson, Department of Botany, Indira Gandhi University, Meerpur Rewari Pin – 122502. The quality, packaging and other particulars of each item should be stated.

WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

1. The quotations are to be sent by registered post in a sealed envelope marked QUOTATION DUE ON 03.02.2021 and this envelope may be put in another envelope addressed to the undersigned.
2. The rates of S.T/C.S.T./GST, EXCISE, CUSTOM and other applicable taxes chargeable must be specified.
3. The rates of Insurance, if any, should be specified. The firm will be required to submit original Payee Receipt along with the bill.
4. Please state the time period within which the items will be supplied.
5. GUARANTEE/WARRANTY PERIOD OF THE INSTRUMENT MUST BE MENTIONED.
6. Discount if any, is to be given with the quotation.
7. The rates quoted should be FOR destination Meerpur, Rewari.
8. Please state, if you are the original manufacturer/authorized distributor of the items quoted by you.

Sr. No	Name of item	Quantity	Make	Model Cat. No.	Other Specification
1.	Orbital Shaker incubator	One	Scigenics(India)	Scigenics LE-TT-A	<ol style="list-style-type: none">1. Speed Range(25 mm Orbit) -25-400 RPM2. Speed Drive –Brushless AC3. Speed Change Via – Soft Touch Keypad4. Temperature Range –Amb + 7° C to 80° C-LT Models 5° C to 80° C LE Models5. Temperature Accuracy² - ± 0.1° C6. Temperature Uniformity³ - ± 0.3° C7. Temperature Sensor – PT 1008. Temperature Control – Microprocessor based PID9. Power Walts – 750W / 1000W10. Electrical Requirements – 6A, 230V,50Hz, 1Ø- LT Models / 16 A, 230V, 50Hz, , 1Ø- LE Models11. Display – Digital LED12. Material of Construction – Exterior – MS Powder Coated / Interior – Stainless Steel13. Weight – 100 Kg (LT) / 12 Kg (LE)14. Tray Size - 420x420mm15. Dimension (W x D x,H) internal – 500 x 490 x 425 (W x D x H) mm16. Dimension (W x D x H) External – 570 x 702 x 815 (W x D x H) mm

Thanking you,

Yours sincerely,

Chairperson 18/01/21

Department of Botany
Chairperson
Indira Gandhi University
Meerpur, Rewari

CC to:-

A copy of the above is forwarded to in-charge University website for uploading the same on University website.