



# INDIRA GANDHI UNIVERSITY, MEERPUR, REWARI

(A State University Established under Haryana Act No. 29 of 2013)

No. IGU/Botany/2020/ 356-370

Dated: 04.12.2020

To

**Subject: - Sealed quotation for Vertical Autoclave NSW-227 (One of 50 liters & One of 98 liters) Due on 21.12.2020.**

Dear Sir/Madam,

Please quote for the article mentioned below so as to reach the undersigned by the Department of Botany. The quality, make, packaging and other particulars of each item should be stated.

WHILE SUBMITTING THE QUOTATIONS PLEASE NOTE THE POINTS MENTIONED BELOW:-

- The quotations are to be sent in a sealed envelope marked QUOTATION DUE ON 21-12-2020 and this envelope may be put in another envelope addressed to the undersigned.
- The rates of S.T/C.S.T./GST, EXCISE, CUSTOM and another taxes chargeable must be specified.
- The rates of Insurance, if any, should be specified. The firm will be required to submit original Payee Receipt along with the bill.
- Please state the time period within which the items will be supplied.
- GUARANTEE PERIOD OF THE INSTRUMENT MUST BE MENTIONED.
- Discount if any, is to be given with the quotation
- The rates quoted should be FOR destination Meerpur, Rewari.
- Please state, if you are the original manufacturer/sole distributor of the items quoted by you.

Sr. No.	Name of item	Quantity	Make	Model Cat. No.	Other Specification
1.	Vertical Autoclave	Two (One of 50 liters & One of 98 liters)	NSW	NSW-227	<ul style="list-style-type: none"><li>• Complete made of stainless steel (304 qty.)</li><li>• Size=350x550(50 Liters) &amp; 450x600( 98 Liters) (Diax Depth)</li><li>• Capacity=50 liters &amp; 98 liters</li><li>• With boiler made of Stainless Steel 16 gauge sheet.</li><li>• Outer shell is also made of Stainless Steel Sheet.</li><li>• Lid is made of Stainless Steel Sheet and tightened around by radial locking.</li><li>• Neoprene rubber gasket.</li><li>• Fitted with water level arrangement.</li><li>• Pressure gauge, Air/Steam release cock two safety valves are also provided.</li><li>• With pizzostat may be set between 10 to 22 PSI.</li><li>• Immersion type heating element.</li><li>• Cut-off automatically when the autoclave is dry.</li><li>• Supplied complete with S.S. basket, cord and plug.</li><li>• Work on 220 Volts 50 cycles A.C. supply.</li><li>• Fully automatic by PID controller.</li></ul>

Thanking you,  
Yours sincerely,

Chairperson

Department of Botany  
Indira Gandhi University  
Meerpur, Rewari

A copy of the above is forwarded to incharge University website for uploading the same on University website.